



POINT-OF-CARE TESTING: CHANGING THE WAY PATIENT CARE IS DELIVERED

MLO and Northern Illinois University (NIU), DeKalb, IL, are co-sponsors in offering continuing education units (CEUs) for this issue's article on **POINT-OF-CARE TESTING: CHANGING THE WAY PATIENT CARE IS DELIVERED**. CEUs or contact hours are granted by the College of Health and Human Sciences at NIU, which has been approved as a provider of continuing education programs in the clinical laboratory sciences by the ASCLS P.A.C.E.® program (Provider No. 0001) and by the American Medical Technologists Institute for Education (Provider No. 121019; Registry No. 0061). Approval as a provider of continuing education programs has been granted by the state of Florida (Provider No. JP0000496), and for licensed clinical laboratory scientists and personnel in the state of California (Provider No. 351). Continuing education credits awarded for successful completion of this test are acceptable for the ASCP Board of Registry Continuing Competence Recognition Program. After reading the article on page 14, answer the following test questions and send your completed test form to NIU along with the nominal fee of \$20. Readers who pass the test successfully (scoring 70% or higher) will receive a certificate for 1 contact hour of P.A.C.E.® credit. Participants should allow four to six weeks for receipt of certificates.

The fee for this continuing education test is \$20.

The Cover Story, Clinical Issues, and Lab Management published in *MLO* are peer-reviewed.

Cover story objectives and the CE questions were prepared by Ellen Olsen, Laboratory Manager, Clinical Laboratory Science Program, Northern Illinois University, DeKalb, IL.

CE QUESTIONS

- The goal of the clinical needs-assessment survey is to**
 - determine which POCT devices are easier to use by the general population.
 - focus on what pathogen clinicians want to test for in a particular disaster scenario.
 - determine if there is a need for wearable or implantable blood-glucose devices for diabetics.
- POCT only impacts hospitals within the healthcare system.**
 - TRUE
 - FALSE
- The goal of POCT is**
 - to promote rapid diagnosis.
 - to provide treatment decisions more quickly.
 - to reduce morbidity and mortality.
 - all of the above.
- The underserved population that is mentioned in this article is**
 - Korea.
 - Africa.
 - India.
- After the tsunami in Southeast Asia and Hurricane Katrina in the United States, it was determined that POCT instruments operated effectively under the harsh environmental conditions of these disasters.**
 - TRUE
 - FALSE
- The workshop between NIBIB and DBT identified which of the following two areas for POCT device development?**
 - Size and durability of devices
 - Low-cost glucose monitoring and platform technologies for multiple diagnostic tests
 - Low-cost glucose and low-cost hemoglobin monitoring
- Of the many POC tests available, which one of the following is *not* mentioned in this article?**
 - Hematocrit
 - Hemoglobin
 - O₂ saturation
 - Blood glucose
- Space shuttles and space stations would not be considered sites where POCT would be utilized.**
 - TRUE
 - FALSE
- Which is *not* true about current POCT devices available in the consumer market and for routine use?**
 - POCT devices meet adequate standards for disaster conditions.
 - POCT devices are compact.
 - POCT devices produce results fast.
- The study done in the Emergency Department in Mexico City by Dr. Mendez evaluated**
 - the TAT in treating patients with chest pain.
 - the cost for the patient using POCT vs. conventional laboratory testing.
 - using POCT to identify cardiac biomarkers.
 - all of the above.
- Which of the following testing devices were *not* able to withstand harsh environmental conditions encountered at disaster sites when tested in a study done by Dr. Louie and his colleagues?**
 - Glucose test strips only
 - Glucose test strips and blood-gas cartridges
 - BUN and creatinine cassettes
 - None of the above.
- Fill in the blanks. Utilization of POCT for evidence-based medical care has the ability to _____ current standards of patient care, while possibly _____ economic costs.**
 - heighten, increasing
 - heighten, decreasing
 - lower, increasing
 - lower, decreasing
- POCT development for various disaster sites and emergency situations has focused on developing pathogen-detection test clusters.**
 - TRUE
 - FALSE

14. The study done by Dr. Tirimacco showed that POCT
- does not facilitate evidence-based medical decisions.
 - decreases the time to diagnosis and treatment of patients.
 - increases cost to the patient.
15. Globally, using POCT will provide the most effective way to diagnose and treat patients in
- high-resource countries.
 - low-resource countries.
16. A critical aspect for efficient POCT device performance is POCT device design.
- TRUE
 - FALSE
17. The use of POCT may
- lower economic costs by facilitating early detection and treatment of patients by doctors and medical personnel.
 - prevent evidence-based medicine to low-resource countries.
 - increase turnaround time.
18. The focus of development for *in vivo* POCT devices which monitor blood chemistry has been for use in
- physician offices.
 - wellness testing areas.
 - hospital and intensive-care units.
19. When designing a POCT device it is not necessary to consider specimen type or volume.
- TRUE
 - FALSE
20. Which of the following is *not* a true statement?
- When selecting a POCT device, the results the device produces should be accurate and precise.
 - When selecting a POCT device, the device should be durable and robust.
 - When selecting a POCT device, it is not necessary to consider if the device meets current requirements for accreditation agencies.
 - When selecting a POCT device, consider if the device is easy to operate and can perform multiple tasks.

MLO's Continuing Education Test is also online.

Both the CE test and a convenient payment feature are available through the auspices of Northern Illinois University.

Go to www.mlo-online.com/CE.aspx to print or to send electronically with payment.

TEST ANSWER FORM

CE Test on POCT: CHANGING THE WAY PATIENT CARE IS DELIVERED

June 2009

(This form may be photocopied; it is no longer valid for CEUs after December 31, 2010)

Circles must be filled in, or test will not be graded.
Shade circles like this: ● Not like this: ⊗

P=Poor; E=Excellent

1. To what extent did the article focus on or clarify the objectives?

P ① ② ③ ④ ⑤ E

2. To what extent was the article well-organized and readable?

P ① ② ③ ④ ⑤ E

3. How will you use the CE units?

- state license
- recertification
- employment
- other

In order to receive CE licensure credits for California or Florida, your license number is required.

Please check applicable box:

- CA Lic. No. _____
- FL Lic. No. _____

FEE FOR THIS CE TEST IS \$20.

Make check payable to:
Northern Illinois University
Fee not refundable or transferable.

Send your check with this form to:
University Outreach Services
Northern Illinois University
DeKalb, IL 60115-2860
Telephone: 815-753-0031

PRINT CLEARLY

NAME _____

MAILING ADDRESS _____ HOME WORK

CITY _____ STATE _____ ZIP _____

INSTITUTION/FACILITY _____

WORK PHONE _____ E-MAIL ADDRESS _____

	A	B	C	D
1.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
2.	<input type="radio"/>	<input type="radio"/>		
3.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
4.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
5.	<input type="radio"/>	<input type="radio"/>		
6.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
7.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
8.	<input type="radio"/>	<input type="radio"/>		
9.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
10.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
11.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
12.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>
13.	<input type="radio"/>	<input type="radio"/>		
14.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
15.	<input type="radio"/>	<input type="radio"/>		
16.	<input type="radio"/>	<input type="radio"/>		
17.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
18.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
19.	<input type="radio"/>	<input type="radio"/>		
20.	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>